

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone (937) 767-7241 (800) 765-4974 Fax (937) 767-9353

Product Model Number:

3163

**Conductivity Calibration
Solution**

Lot Number:

10A100023

Testing of Production Standard with a Metrology Calibrated Cell

**Nominal Conductivity
(microsiemens/cm):**

10K

**Specification Range,
(microsiemens/cm):**

9,900 - 10,100

**Test Result,
(microsiemens/cm):**

9985

Lot Result:

Accepted

Inspected by::

J. Jenkins

Date Tested:

05-Jan-10



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility