

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category                      Buffer and solid products  
Product Model Number                2392  
Lot Number                                18A100674  
Method                                      Verification of mass

Determination	Specification	Result
Average Mass	0.20-0.30 grams	0.26 g

Conforms to acceptance criteria

Completed by: Kristin Stover  
Your Name

Date: 1/25/2018



A handwritten signature in blue ink, appearing to read 'W. Mabel'.

Director, Quality and Corporate  
Responsibility