

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2365
Lot Number 19A100356
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|---------------|----------|
| Initial | Average Calibration Sensitivity | 8.0-45.0 nA | 11.0 nA |
| | Average Linearity | 8.55-9.45 g/L | 8.99 g/L |
| | Potassium Ferrocyanide | 0.00-0.05 | 0.00 g/L |

Conforms to acceptance criteria

Completed by: Sandy Crawford
 Your Name

1/24/2019



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility