

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category                    Bioanalytical Standard  
Product Model Number            2778 Sucrose Std  
Lot Number                            19D100586  
Method                                Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Sucrose	(Average of production) - (average of metrology)	$\pm 0.3$ g/L	0.05 g/L
Sucrose	Coefficient of variation for production standard (CV, %)	$\leq 2\%$	0.4%
N/A	(Average of production) - (average of metrology)	$\pm$ N/A	N/A
N/A	Coefficient of variation for production standard (CV, %)	$\leq$ N/A	N/A

Conforms to acceptance criteria

Completed by: Heather Brooks

Date: 4/24/2019

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



Gregory W. Popp  
Quality Manager, Analytics