

# Certificate of Analysis



YSI Incorporated  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Bioanalytical Standard  
Product Model Number 2726  
Lot Number 20B130082  
Method Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Methanol A, 1.00 g/L	(Average of production) - (average of metrology)	$\pm 20$ mg/L	4.4 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	0.90%
Methanol B, 2.50 g/L	(Average of production) - (average of metrology)	$\pm 50$ mg/L	12.2 mg/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	1.1%

Conforms to acceptance criteria

Completed by: David Stoler Date: 3-3-2020

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



Gregory W. Popp  
Quality Manager, Analytics