

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Sucrose Membrane
Product Model Number 2703
Lot Number 21G100431
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|-------------------------------------|-----------------|-----------|
| Initial | Average Calibration Sensitivity | 10.0-40.0 nA | 11.1 nA |
| | Average Linearity | 23.75-26.25 g/L | 24.67 g/L |
| | Potassium Ferrocyanide Interference | 0.0-0.3 | 0.00 g/L |

Conforms to acceptance criteria

Completed by: Christy Woodford
Your Name

Date: 7/20/2021

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics