

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2329
Lot Number 23A100286
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|------------------|--------------|
| Initial | Average Calibration Sensitivity | 6.0-30.0 nA | 15.5 nA |
| | Average Linearity | 14.3-15.7mmol/L | 14.92 mmol/L |
| | Potassium Ferrocyanide | -0.3 - 0.3mmol/L | 0.00 mmol/L |

Conforms to acceptance criteria

Completed by: Christy Woodford 1/17/2023
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics