

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2735  
Lot Number                 23K100006  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	11.2 nA
	Average Linearity	7.6-8.4mmol/L	7.83 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Sandy Crawford  
Your Name

10/2/2023



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics