

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Membrane
Product Model Number	2771
Lot Number	23M100144
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	>5 nA	nA
	Average Linearity	427 - 473 mg/L	mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	mg/L
Day 2	Average Calibration Sensitivity	>5 nA	nA
	Average Linearity	427 - 473 mg/L	mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	mg/L

Conforms to acceptance criteria

Completed by:

Christy Woodford

12/13/2023

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager Analytics