

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2776  
Lot Number 24A100603  
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination   | Tolerance            | Result   |
|------------------|---|----------------------|----------|
| GLUCOSE          | (Average of Production)-(Average of Metrology)          | $\pm 0.03\text{g/L}$ | 0.02 g/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$         | 0.5 %    |
| LACTATE          | (Average of Production)-(Average of Metrology)          | $\pm 0.01\text{g/L}$ | 0.00 g/L |
|                  | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$         | 0.5 %    |

Conforms to acceptance criteria

Completed by:

Nick Rittenhouse

Date: 1/31/2024

1/31/2024

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



Gregory W. Popp  
Quality Manager, Analytics