

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2702
Lot Number 24B100517
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 30 nA	14.86 nA
	Average Linearity	23.75 – 26.25 g/L	25.0 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = <u>N/A</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by: Christy Woodford *Christy Woodford* Date: 2/27/2024



Lisa M. Abel
Lisa M. Abel
Director, Quality and Corporate
Responsibility