

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353


Product Category Membrane
Product Model Number 2702
Lot Number 24C100146
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 30 nA	13.83 nA
	Average Linearity	23.75 – 26.25 g/L	24.76 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = <u>N/A</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by: Christy Woodford  Date: 3/13/2024




Lisa M. Abel
Director, Quality and Corporate
Responsibility