

Certificate of Analysis




YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353


Product Category Membrane
Product Model Number 7140
Lot Number 24C100449
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Result |
|------------------|-------------------------------------|---------------|----------|
| Initial | Average calibration sensitivity | 10 – 75 nA | 31.8 nA |
| | Average Linearity | 38 – 42 g/L | 38.8 g/L |
| | Potassium Ferrocyanide interference | N/A | N/A |
| Day = <u>N/A</u> | Average calibration sensitivity | N/A | N/A |
| | Average Linearity | N/A | N/A |
| | Potassium Ferrocyanide interference | N/A | N/A |

Conforms to acceptance criteria

Completed by: Christy Woodford  Date: 3/27/2024




Lisa M. Abel
Director, Quality and Corporate
Responsibility