

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2701
Lot Number 24C100451
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 50 nA	12.12 nA
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	+/-1.0 nA	0.028nA
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by: Sandy Crawford

Date: 3/26/2024

Sandy Crawford



Ronald A. Johnson

Ron A. Johnson
Director of Quality Systems