

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353


Product Category            Membrane  
Product Model Number      2702  
Lot Number                    24D100109  
Method                         Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 30 nA	17.25 nA
	Average Linearity	23.75 – 26.25 g/L	24.62 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = <u>N/A</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by: Christy Woodford  Date: 04/11/2024



  
Lisa M. Abel  
Director, Quality and Corporate  
Responsibility