

# Technical Compliance Statement



**For the following information**

Ref. File No.: ACI-2011266

Product : Portable Conductivity, Salinity and Temperature Instrument  
Model Number : CO310, EC300A  
Applicant : SHANGHAI JENCO INSTRUMENTS CO., LTD  
Manufacturer : SHANGHAI JENCO INSTRUMENTS CO., LTD  
Standards :

EN 61326-1:2006 (IEC 61326-1:2005) (Basic immunity test requirements)  
(IEC 61000-4-2:2001, IEC 61000-4-3:2006,  
IEC 61000-4-4:2004+Corr.1:2006+Corr.2:2007, IEC 61000-4-6:2006)

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 2004/108/EC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. The test data & results are issued on the EMC test report no. ACI-E11116.

Signature

Byron Kwo/Senior Manager

Date: Aug 31, 2011

Test Laboratory:  
Audix Technology (Shanghai) Co., Ltd.  
NVLAP Lab Code: 200371-0  
Web Site: [www.audixtech.com](http://www.audixtech.com)

**NVLAP**<sup>®</sup>  
Lab Code: 200371-0

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.