

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2329
Lot Number 17A100099
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-30.0 nA	20.7 nA
	Average Linearity	14.3-15.7mmol/L	14.94 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Nicholas Hetherington 2/6/17
 Your Name



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility