

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
 Product Model Number 2365
 Lot Number 17E100281
 Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	10.5 nA
	Average Linearity	8.55-9.45 g/L	8.89 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by: Amanda Mills

 Your Name

5/19/2017



Lisa M. Abel
 Director, Quality and Corporate
 Responsibility