

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2754  
Lot Number                 19D100590  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	09.5 nA
	Average Linearity	9.50-10.50mmol/L	9.91 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.02 mmol/L

Conforms to acceptance criteria

Completed by:            Sandy Crawford  
                                  Your Name

4/24/2019



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility