

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products
Product Model Number 2357
Lot Number 200212GFS098005
Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.35g \pm 0.35g	6.27g
Average pH (buffers only)	7.3 \pm 0.2	7.2

Conforms to acceptance criteria

Completed by: David Stoler

Date: 3-11-2020

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp
Quality Manager, Analytics