

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Bioanalytical Standard
Product Model Number 2767
Lot Number 20A100485
Method Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Dextrose	(Average of production) - (average of metrology)	± 1.0 g/L	0.2 g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	1.44%

Conforms to acceptance criteria

Completed by: David Stoler

Date: 1-23-2020

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics