

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Buffer and solid products
Product Model Number	2363
Lot Number	20B130022
Method	Verification of mass

---

Determination	Specification	Result
Average Mass	1.15-1.35 grams	1.26 g

---

Conforms to acceptance criteria

Completed by: Lucas Stover

Date: 2/20/2020



A handwritten signature in blue ink, appearing to read 'W. Mabel'.

Director, Quality and Corporate  
Responsibility