

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard  
Product Model Number 2778  
Lot Number 20D100369  
Method Comparative enzymatic testing of Production standard vs. Metrology standard

<b>Chemistry Tested</b>	<b>Determination</b>	<b>Tolerance</b>	<b>Result</b>
SUCROSE	(Average of production) - (average of metrology)	$\pm 0.3$ g/L	0.0 g/L
	Coefficient of variation for production standard (CV, %)	$\leq 2$ %	0.5%
	(Average of production) - (average of metrology)	N/A	N/A
	Coefficient of variation for production standard (CV, %)	N/A	N/A

Conforms to acceptance criteria

Completed by: Lucas Stover Date: 04/18/2020



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility