

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Membrane
Product Model Number	2771
Lot Number	22J100374
Method	Amperometric testing with standards

---

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	>5 nA	12.62nA
	Average Linearity	427 - 473 mg/L	452.00mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.00mg/L
Day 2	Average Calibration Sensitivity	>5 nA	11.48nA
	Average Linearity	427 - 473 mg/L	435.60mg/L
	Potassium Ferrocyanide Interference	<15 mg/L	0.01mg/L

---

Conforms to acceptance criteria

Completed by:

Christy Woodford

9/27/2022

Your Name



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics