

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Check Solution
Product Model Number 7173
Lot Number 23K100054
Method Comparative conductivity testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Sodium	(Average of production) - (average of metrology)	$\pm 1.0\%$	0.71%
	Coefficient of variation for production standard (CV, %)	$\leq 2.0\%$	0.52%

Conforms to acceptance criteria

Completed by: Amanda Fenner

Date: 10/07/2023

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics