

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Sucrose Membrane
Product Model Number	2703
Lot Number	23L100669
Method	Amperometric testing with standards

---

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-40.0 nA	09.7 nA
	Average Linearity	23.75-26.25 g/L	25.18 g/L
	Potassium Ferrocyanide Interference	0.0-0.3	0.01 g/L

---

Conforms to acceptance criteria

Completed by: \_\_\_\_\_ Christy Woodford

Date: 12/6/2023

Your Name

*The procedures used to manufacture and test this product have been verified by QA for GMP compliance.*



Gregory W. Popp

Quality Manager Analytics