

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2761
Lot Number 24B100405
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 – 60 nA	43.88 nA
	Average Linearity	27.0 – 33.0 g/L	29.06 g/L
	Potassium Ferrocyanide interference	0 – 0.07 g/L	0.00 g/L
Day = <u>N/A</u>	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

Conforms to acceptance criteria

Completed by:

Christy Woodford
Christy Woodford

Date:

2/21/2024



Lisa M. Abel

Lisa M. Abel
Director, Quality and Corporate
Responsibility