



**Audix Technology (Shanghai) Co., Ltd.**

3F 34Bldg 680 Guiping Rd.,  
Caohejing Hi-Tech Park, Shanghai, China 200233

Tel: +86-21-64955500  
Fax: +86-21-64955491  
E-Mail: [audixaci@audix.com](mailto:audixaci@audix.com)

No.: ACI-2008064

# Statement of Compliance

Regarding the statement of products which are in the scope of  
the Council Directive 2004/108/EC

Shanghai Jenco Instruments Co., Ltd.

18 Wangdongzhong Road, Sijing Town, Songjiang, Shanghai, China.

has successfully demonstrated that its product

**Dissolved Oxygen Meter**

**Model No. : PRO20**


**Serial No. : E08032506**


is in compliance with

EN 61326:1997+A1:1998+A2:2001+A3:2003  
(IEC 61326:1997+A1:1998+A2:2000+A3:2002)  
(IEC 61000-4-2:2001, IEC 61000-4-3:2006,  
IEC 61000-4-4:2004+Corr.1:2006+Corr.2:2007, IEC 61000-4-6:2006)

as described in the Test Report ACI-E08046.

(Date: May 12, 2008)

 For and on behalf of  
Audix Technology (Shanghai) Co., Ltd.

  
Authorized Signature EMC **BYRON KWO**  
Manager



The statement is based on a single evaluation of samples of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.