

ADMINISTRATIVE DATA

Manufacturer:

YSI
1700/1725 Brannum Ln
Yellow Springs, OH 45387

Appliance/Product: ProODO Water Quality Measurement Instrument

Model/Type Number: 626281/626284-xxx

Rating: 3VDC @0.25A

RF Suppression Components: see attached sheets.

Measurement Equipment used: see attached sheets.

Measurements According to, and Sample Unit Complies with:

FCC 47 CFR Part 15-2008 Class B, using ANSI C63.4:2003
ICES-003:2004 Class B
EN 61326-1:2006 Class B
EN 61326-2-3:2006
EN 61000-3-2:2006 Class A
EN 61000-3-3:1995 +A1:2001+A2:2005
AS/NZS 4251.1:1999

Report Prepared By: John R. Barnes KS4GL, PE, NCE, NCT, ESDC Eng, ESDC Tech, PSE,
SM IEEE

Testing Performed by:


dBi Corporation
216 Hillsboro Avenue
Lexington, KY 40511-2105, USA

on November 19-21, 2008

at: Lexmark International, Inc.
Development Lab.
Lexington, KY 40550, USA

Reviewed and Approved by: John R. Barnes, KS4GL, PE, NCE, NCT, ESDC Eng, ESDC
Tech, PSE, SM IEEE

SIGNED


John R. Barnes, PRESIDENT dBi Corp.

Date: November 22, 2008