Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2365

Lot Number 16M100566

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
	Average Calibration Sensitivity	8.0-45.0 nA	14.2 nA
	Average Linearity	8.55-9.45 g/L	8.93 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by: Nic Hetherington 1/5/2017

Christy Woodford



Lisa M. Abel Director, Quality and Corporate Responsibility