Certificate of Analysis

YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product CategoryMembraneProduct Model Number2365Lot Number17A100587MethodAmperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-45.0 nA	15.3 nA
	Average Linearity	8.55-9.45 g/L	8.85 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by:

Nicholas Hetherington Your Name 2/14/2017



n Malel

Lisa M. Abel Director, Quality and Corporate Responsibility