Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 17F100252

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
	Average Calibration Sensitivity	10.0-30.0 nA	24.1 nA
	Average Linearity	14.3-15.7mmol/L	14.96 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	mmol/L

Conforms to acceptance criteria

Completed by: Amanda Mills

Your Name



Lisa M. Abel
Director, Quality and Corporate
Responsibility