Certificate of Analysis

YSI Incorporated

YSI

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product CategoryMembraneProduct Model Number2329Lot Number17H100291MethodAmperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10.0-30.0 nA	19.8 nA
	Average Linearity	14.3-15.7mmol/L	14.75 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by:

Amanda Mills 08/08/2017

Your Name



malel

Lisa M. Abel Director, Quality and Corporate Responsibility