## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 18A100111

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results	
Initial	Average Calibration Sensitivity	6.0-30.0 nA	18.1 nA	
	Average Linearity	14.3-15.7mmol/L	14.84 mmol/L	
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.00 mmol/L	

/ `~ ~ t	- C PPO C	+~	acceptance	~ri+~	
1 ./ 1/ 1/		111	arremance	1 'I II <u>~</u>	11111

Completed by: Amanda Mills 01/10/2018

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Lisa M. Abel
Director, Quality and Corporate
Responsibility