## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 18A100678

Method Amperometric testing with standards

| Time of Testing | Determination                   | Specification    | Results      |
|-----------------|---------------------------------|------------------|--------------|
| Initial         | Average Calibration Sensitivity | 6.0-30.0 nA      | 16.3 nA      |
|                 | Average Linearity               | 14.3-15.7mmol/L  | 14.94 mmol/L |
|                 | Potassium Ferrocyanide          | -0.3 - 0.3mmol/L | 0.00 mmol/L  |

| ~ ′        |          |               |
|------------|----------|---------------|
| ( `ontormo |          | ance criteria |
| COULOURS   | o acceor | ance cinena   |
|            |          |               |

Completed by: Amanda Mills 01/30/2018

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Lisa M. Abel
Director, Quality and Corporate
Responsibility