## **Certificate of Analysis**



## **YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2365

Lot Number 18J100105

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
	Average Calibration Sensitivity	8.0-45.0 nA	10.2 nA
	Average Linearity	8.55-9.45 g/L	9.11 g/L
	Potassium Ferrocyanide	0.00-0.05	0.00 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford/Sandy Crawford

Your Name



Lisa M. Abel Director, Quality and Corporate Responsibility

mullel

9/12/2018