Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2365

Lot Number 18J100699

Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|---------------|----------|
| | Average Calibration Sensitivity | 8.0-45.0 nA | 11.0 nA |
| | Average Linearity | 8.55-9.45 g/L | 9.05 g/L |
| | Potassium Ferrocyanide | 0.00-0.05 | 0.00 g/L |

Conforms to acceptance criteria

Completed by: Sandy Crawford 10/4/2018

Your Name



Lisa M. Abel Director, Quality and Corporate Responsibility

mullel