Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2329

Lot Number 18K100141

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	17.4 nA
	Average Linearity	14.3-15.7mmol/L	14.68 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.02 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 10/9/2018

Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Lisa M. Abel
Director, Quality and Corporate
Responsibility