## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Membrane

Product Model Number 2754

Lot Number 18M100211

Method Amperometric testing with standards

| Time of Testing | Determination                   | Specification    | Results     |
|-----------------|---------------------------------|------------------|-------------|
| Initial         | Average Calibration Sensitivity | 8.0-40.0 nA      | 10.1 nA     |
|                 | Average Linearity               | 9.50-10.50mmol/L | 9.95 mmol/L |
|                 | Potassium Ferrocyanide          | < 0.4mmol/L      | 0.05 mmol/L |

Conforms to acceptance criteria

Completed by: Christy Woodford 12/19/2018

Your Name



Lisa M. Abel Director, Quality and Corporate Responsibility