Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and solid products
Product Model Number	2357 Buffer Kit
Lot Number	190524GFS098005
Method	Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00-6.70 g	6.18G
Average pH (buffers only)	7.1-7.5	7.2 pH

Conforms to acceptance criteria					
Completed by:	Lucas Stover	Date:	06/06/19		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics