Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Buffer and solid products

Product Model Number 2357 Buffer Kit

Lot Number 190808GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00-6.70 g	6.16 g
Average pH (buffers only)	7.1-7.5	7.28 pH

Conforms to acceptance criteria					
Completed by:	David Stoler	Date:	09/13/2019		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp