Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category Buffer and solid products

Product Model Number 2357

Lot Number 190916GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	6.00 - 6.70 g	6.25g
Average pH (buffers only)	7.1-7.5	7.3

Completed by: Lucas Stover Date: 10/16/2019



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel