Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Buffer and so	alid products
Product Category	Durier and so	ma products

Product Model Number 2357

Lot Number 191106GFS098005

Method Verification of mass and pH (for buffers)

Determination	Specification	Result
Average mass	$6.35 \pm .35 \text{ g}$	6.17 g
Average pH (buffers only)	$7.3 \pm .2$	7.2

Conforms to accept	tance criteria			
Completed by:	David Stoler	Date:	12-9-2019	

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp