Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category ISE Standard

Product Model Number 2972 Ammonium Standard

Lot Number 19B101081

Method Comparative ISE testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of production) - (average of	< 0.250 mV	0.077 mV
	metrology)		
Ammonium	Coefficient of variation for production	≤ 2%	0.19%
	standard (CV, %)		
Potassium	(Average of production) - (average of	< 0.250 mV	0.21 mV
	metrology)		
Potassium	Coefficient of variation for production	≤ 2%	0.525
	standard (CV, %)		

Conforms to acceptance criteria					
Completed by:	Christina M. Kidder	Date:	03/06/2019		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp