Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category ISE Standard

Product Model Number 2972 Ammonium Std

Lot Number 19E100157

Method Comparative ISE testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of production) - (average of metrology)	<0.250mV	0.034mV
Ammonium	Coefficient of variation for production standard (CV, %)	≤2%	0.14%
Potassium	(Average of production) - (average of metrology)	<0.250mV	0.157mV
Potassium	Coefficient of variation for production standard (CV, %)	≤2%	0.40%

Conforms to acceptance criteria						
Completed by:	Heather Brooks		Date:	5/16/19		

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Gregory W. Popp Quality Manager, Analytics

Dregory W. Popp