Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category	Bioanalytical Standard
Product Model Number	2767 Xylose
Lot Number	<u>19E100930</u>
Method	Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Xylose	(Average of production) - (average of metrology)	±1.0g/L	0.24g/L
Xylose	Coefficient of variation for production standard (CV, %)	≤2.0%	1.78%
N/A	(Average of production) - (average of metrology)	\pm N/A	N/A
N/A	Coefficient of variation for production standard (CV, %)	\leq N/A	N/A

Conforms to acceptance criteria

Completed by: Heather Brooks

Date: 5/30/2019

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Dregory W. Popp

Gregory W. Popp Quality Manager, Analytics