Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	ISE Standard	
Product Model Number	7179	
Lot Number	119F100136	
Method	Comparative ISE testing of Production standard vs. Metrology stand	ard

Chemistry Tested	Determination	Tolerance	Result
AMMONIUM	(Average of production) - (average of metrology)	<0.250 mV	0.228 mV
	Coefficient of variation for production standard (CV, %)	≤2%	0.10%
POTASSIUM	(Average of production) - (average of metrology)	<0.250 mV	0.027 mV
	Coefficient of variation for production standard (CV, %)	≤2%	0.16%

Conforms to acceptance criteria						
Completed by:	Lucas Stover		Date:	06/26/2019		



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel