## **Certificate of Analysis**



**YSI Incorporated** 

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2726	_
Lot Number	19F100709	_
Method	Comparative enzymatic testing	g of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
Methanol Std A	(Average of production) - (average of	± 20 mg/L	6.6 mg/L
1.0 g/L	metrology)		
	Coefficient of variation for production	≤ 2 %	1.1%
	standard (CV, %)		
Methanol Std B	(Average of production) - (average of	$\pm$ 50 mg/L	1.1 mg/L
2.50 g/L	metrology)		
	Coefficient of variation for production	≤ 2 %	0.7%
	standard (CV, %)		

Conforms to acceptance criteria					
Completed by:	Lucas Stover	Date:	10/02/2019		



Lisa M. Abel Director, Quality and Corporate Responsibility

Turalel