Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

| Product Category | ISE Standard | |
|----------------------|----------------------------|--|
| Product Model Number | 7179 | <u></u> |
| Lot Number | 19G100132 | <u></u> |
| Method | Comparative ISE testing of | Production standard vs. Metrology standard |

| Chemistry Tested | Determination | Tolerance | Result |
|-------------------------|--|-----------|----------|
| AMMONIUM | (Average of production) - (average of metrology) | <0.250 mV | 0.246 mV |
| | Coefficient of variation for production standard (CV, %) | ≤2% | 0.11% |
| POTASSIUM | (Average of production) - (average of metrology) | <0.250 mV | 0.183 mV |
| | Coefficient of variation for production standard (CV, %) | ≤2% | 0.13% |

| Conforms to acce | ptance criteria | | |
|------------------|-----------------|-------|------------|
| Completed by: | Lucas Stover | Date: | 07/17/2019 |



Lisa M. Abel Director, Quality and Corporate Responsibility

Thralel