Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Bioanalytical Standard	
Product Model Number	2792	_
Lot Number	19G100360	_
Method	Comparative enzymatic testir	g of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 1.0G/L	(Average of production) - (average of metrology)	±20mg/L	5.0 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.5%
ETHANOL STD B 0.5G/L	(Average of production) - (average of metrology)	±10mg/L	6.5 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.6%

Conforms to acceptance criteria					
Completed by:	Lucas Stover	Date:	07/19/2019		



Lisa M. Abel Director, Quality and Corporate Responsibility

Tomalel